

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

## CB TEST CERTIFICATE

Product

Switching Mode Power Supply

Name and address of the applicant

MEAN WELL Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248,  
Taiwan

Name and address of the manufacturer

MEAN WELL Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248,  
Taiwan

Name and address of the factory

Note: When more than one factory, please report on page 2

MEAN WELL Enterprises Co., Ltd.  
No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248  
Taiwan Additional Information on page 2

Ratings and principal characteristics

See Page 2

Trademark (if any)



Type of Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

MFM-x-y, MPM-x-y  
See Page 2

Additional information (if necessary may also be reported on page 2)

Additionally evaluated to EN 60601-1:2006/ A1:2013/ A12:2014;  
National Differences specified in the CB Test Report. Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

As shown in the Test Report Ref. No. which forms part of this Certificate

1706036-CB issued on 2017-10-25

This CB Test Certificate is issued by the National Certification Body



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2017-10-30

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

**DK-67656-UL**

**Model Details:**

MFM-x-y,MPM-x-y (x can be 15 or 20, y can be 3.3, 5, 12, 15 or 24)

**Factories:**

SuZhou MEAN WELL Technology Co., Ltd.

No. 77, Jian-min Road, Dong-qiao, Pan-yang Ind. Park, Huang-dai Town, Xiang-cheng District, Suzhou, Jiangsu 215152

P.R. China

MEAN WELL (GUANGZHOU) ELECTRONICS CO., LTD HUADU BRANCH.

No.11 Jingu South Road, Huadong Town, Huadu District, Guangzhou China

**Ratings:**

For Model MFM-15-3.3, MFM-15-5, MFM-15-12, MFM-15-15, MFM-15-24:  
100-240VAC, 0.6A-0.3A, 50/60Hz

For Model MFM-20-3.3, MFM-20-5, MFM-20-12, MFM-20-15, MFM-20-24:  
100-240VAC, 0.75A-0.5A, 50/60Hz

See test report for details.

**Additional Information:**

Risk Management was not included in this investigation.

**Additional information (if necessary)**



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2017-10-30

Signature:

Jan-Erik Storgaard